

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0925340	<b>(X3) Date Survey Completed</b>  10/27/2020
<b>Name of Provider or Supplier</b>  Jeffrey L Stein Md Pa	<b>Street Address, City, State</b>  9291 Glades Rd Ste 306, Boca Raton, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at Jeffrey L Stein MD PA on October 27, 2020. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
<b>D2121</b>	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of American Proficiency Institute (API) proficiency testing results and interview with the Technical Consultant, the laboratory failed to score at least 80% for Hematocrit in 1 out of 6 (3rd testing event 2018 through 2nd testing event in 2020) testing events reviewed in the specialty of Hematology. Findings Included: Review of API proficiency testing results in the 2nd testing event in 2020 revealed a score of 20 % for Hematocrit. During an interview on October 27, 2020 at 10:30 AM, the Technical Consultant confirmed the proficiency testing failure.</p>