

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0927895	<b>(X3) Date Survey Completed</b>  03/23/2021
<b>Name of Provider or Supplier</b>  Skin Cancer Treatment Center	<b>Street Address, City, State</b>  10001 Pines Blvd, Pembroke Pines, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A recertification survey was completed on 3/23/2021 at Skin Cancer Treatment Center. The laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following conditions were not met : 5400- Analytic Systems
<b>D5400</b>	<p>ANALYTIC SYSTEMS CFR(s): 493.1250</p> <p>Each laboratory that performs nonwaived testing must meet the applicable analytic systems requirements in 493.1251 through 493.1283, unless HHS approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub.7), that provides equivalent quality testing. The laboratory must monitor and evaluate the overall quality of the analytic systems and correct identified problems as specified in 493.1289 for each specialty and subspecialty of testing performed.</p> <p>This CONDITION is not met as evidenced by: Based on observation and interview, the laboratory failed to have Histology reagents that were not expired in the chemical cabinets.(See 5417)</p>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview, the laboratory failed to have Histology reagents that were not expired in the chemical cabinets. Findings Included: This is a repeated deficiency from 4/4/2019. During an observation of the chemical cabinets the</p>

following Histology reagents were revealed as expired: 1. Eosin Working Solution expired 10/16/19 2. OCT Embedding Matrix for Frozen Sections expired 01/2021 3. Submount mounting medium expired 01/02/2021 4. SafeClear Xylene Substitute expired 08/2020 During an interview on 3/23/2020 at 10:00am, the laboratory director confirmed the histology reagents were expired.