

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0928464	(X3) Date Survey Completed 03/06/2023
Name of Provider or Supplier Terrence A Cronin Jr Md	Street Address, City, State 1399 S Harbor City Blvd, Melbourne, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Rcertification survey was conducted on March 6, 2023. Terrence A Cronin Jr MD clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5609	<p>HISTOPATHOLOGY CFR(s): 493.1273(e)(f)</p> <p>(e) The laboratory must use acceptable terminology of a recognized system of disease nomenclature in reporting results. (f) The laboratory must document all control procedures performed, as specified in this section.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to document quality control (QC) information including open dates for all reagents used in their Hematoxylin & Eosin (H & E), from March 2021 to April 2022. Findings: The laboratory used the following reagents in their H&E stain: Scott Tap Water, 100% Reagent Alcohol, Gill's Hematoxylin III, and Eosin Working Solution. Review of the laboratory's procedure manual noted for 493.1273" QC records should include lot numbers, date opened/prepared, expiration dates . . . " Review of the laboratory's QC logs titled "Reagent Control Sheet" showed there were no open dates for reagents used prior to the reagents currently being used. On 03/06/2023 at 10:50 AM, Mohs Technician B acknowledged the open dates were not recorded.</p>