

| | | |
|--|---|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 10D0929531 | (X3) Date Survey Completed 05/15/2019 |
| Name of Provider or Supplier Boca Raton Physicians Pa | Street Address, City, State 1905 Clint Moore Rd Ste 201, Boca Raton, FL | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D2087 | <p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory received an unsatisfactory result in routine chemistry. Findings include: During review of the past two years of proficiency testing on 05/15/2019, the surveyor found that the laboratory received an unsatisfactory result of 0% for chloride. During an interview with the testing person at 12:55 p.m. on 05/15/2019, he confirmed that they had received an unsatisfactory score.</p> |
| D2099 | <p>ENDOCRINOLOGY CFR(s): 493.843(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, the laboratory received an unsatisfactory result in endocrinology. Findings include: During review of the past two years of proficiency testing on 05/15/2019, the surveyor found that the laboratory received an unsatisfactory result of 40% for triiodothyronine. During an interview with the testing person at 12:55 p.m. on 05/15/2019, he confirmed that they had received an unsatisfactory score.</p> |
| D2122 | HEMATOLOGY |

CFR(s): 493.851(b)

Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.

This STANDARD is not met as evidenced by:

Based on record review and interview with laboratory personnel, the laboratory received an unsatisfactory result in hematology. Findings include: During review of the past two years of proficiency testing on 05/15/2019, the surveyor found that the laboratory received an unsatisfactory result of 73% for cell identification or white blood cell identification. During an interview with the testing person at 12:55 p.m. on 05/15/2019, he confirmed that they had received an unsatisfactory score.