

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0936067	(X3) Date Survey Completed 09/24/2019
Name of Provider or Supplier Associates In Dermatology Inc	Street Address, City, State 530 Ocoee Commerce Pkwy, Ocoee, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Recertification survey was conducted on September 24, 2019. Associates in Dermatology Inc. clinical laboratory was found not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview, the laboratory used an expired reagent in their Hematoxylin and Eosin (H&E) stain from 4/19/19 to 9/24/19. Findings: During a tour of the histopathology laboratory on 9/24/19 at 9:10 AM an expired bottle of Scott's Tap Water Substitute lot number 1810305 was located in their flammable cabinet. The expiration date on the bottle of Scott's Tap Water Substitute was listed as 4/18/19. The new bottle of Scott's Tap Water Substitute lot number 1916501 (expiration 6/19/20) in the flammable cabinet and recorded on the laboratory's "Chemical Inventory Log" didn't list an open date. During an interview on 9/24/18 at 9:10 PM, the Operations Manager acknowledged that the Scott's Tap Water Substitute was expired and was being used in their H&E stain.</p>