

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0938578	<b>(X3) Date Survey Completed</b>  04/16/2019
<b>Name of Provider or Supplier</b>  Key Biscayne Pediatric Associates Inc	<b>Street Address, City, State</b>  240 Crandon Blvd #212, Key Biscayne, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2121</b>	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of American Proficiency Institute (API) proficiency testing records and interview with Office Manager (OM), the laboratory failed to score at least 80 % on Blood Cell Identification (BCI) for 1 (3rd event of 2018) out of 6 events for Hematology reviewed. Findings include: Review of API proficiency records revealed a score of 0% for BCI in the 3rd event of 2018. During an interview on 04/16 /2019 at 11:30 AM, the OM confirmed the failed analyte proficiency testing score.</p>
<b>D2122</b>	<p>HEMATOLOGY CFR(s): 493.851(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on record review of American Proficiency Institute (API) proficiency testing records and interview with Office Manager (OM), the laboratory received an unsatisfactory score in 1 (2rd event of 2017) out of 6 events for Hematology specialty reviewed. Findings include: Review of API proficiency records for 2017 and 2018 revealed that the laboratory received a 20 % score for Red Blood Cells, Hematocrit and Hemoglobin resulting in an overall score of 60 % during the second event of 2017. During an interview on 04/16/2019 at 11:30 AM, the OM confirmed that the laboratory failed the event of reference.</p>