

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0943459	(X3) Date Survey Completed 05/15/2019
Name of Provider or Supplier Abrams Dermatology Llc	Street Address, City, State 3328 Bee Ridge Rd, Sarasota, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Abrams Dermatology LLC on 05/15/19. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with Testing Person #B the laboratory failed to assess competency on 2 (Testing Person #B and #C) out of 3 Testing Personnel (#A, #B, and #C) for 2 out of 2 (2017-2019) years reviewed. Findings Included: Review of the CMS 209, Laboratory Personnel Report, signed by the Laboratory Director on 05/06/19 revealed 3 moderate complexity Testing Personnel, #A, #B, and #C. Review of employee competency evaluations revealed no competencies on Testing Person #B or #C. An interview on 05/15/19 at 11:54 AM with Testing Person #B revealed that there was no policy to perform the competency evaluations and confirmed that the competencies on Testing Person #B and #C had not been conducted.</p>
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p>

This STANDARD is not met as evidenced by:

Based on record review and interview with Testing Person #B the laboratory failed to verify the accuracy of testing at least twice a year for 1 (2018) out of 2 years (2017-2019) reviewed for scabies testing in the subspecialty of Parasitology. Findings Included: Review of the peer review testing revealed that it had only been performed December of 2018 (12/18) for scabies testing in 2018. Interview with Testing Person #B on 05/15/19 at 11:48 AM confirmed the laboratory only verified the accuracy of scabies testing once in 2018.