Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 10D0943459	**(X3) Date Survey Completed** 08/11/2021
Name of Provider or Supplier Abrams Dermatology Llc	**Street Address, City, State** 3328 Bee Ridge Rd, Sarasota, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	An announced CLIA recertification survey was conducted at Abrams Dermatology LLC on 08/11/2021. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on record review and interview with the Laboratory Director, the laboratory failed to verify the accuracy of testing twice a year for 2 (2019 -2020) of 2 years (2019-2020) reviewed for Parasitology. Findings Included: Review of Quality Assurance form revealed no peer reviews for 2019 and 1 peer review (January 2020) was completed in 2020 for for Scabies testing in Parasitology. Interview on 08/11/21 at 10:50 AM, the Laboratory Director confirmed that the laboratory had only performed peer review once, which was in 2020. This is a repeat deficiency.