

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0953693	<b>(X3) Date Survey Completed</b>  07/25/2018
<b>Name of Provider or Supplier</b>  Genesis Dermatology	<b>Street Address, City, State</b>  600 Heritage Dr Ste 101, Jupiter, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review, and interview with laboratory personnel, there was an expired reagent in the flammable cabinet in the laboratory. Findings include: During tour of the laboratory at 11:30 a.m. on 07/25/2018, the surveyor found a bottle of Gill 3 Hematoxylin that expired on 04/18. The bottle was received on 01/18/2018 and opened on 01/28/18. The surveyor reviewed the reagent logs, and this lot of hematoxylin was not logged in. During an interview with the office manager at 12:05 p.m. on 07/25/2018, she said that it was probably the bottle that a neighboring laboratory borrowed when they ran out, and maybe that was why it had not been logged in.</p>