

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0953693	<b>(X3) Date Survey Completed</b>  09/02/2020
<b>Name of Provider or Supplier</b>  Genesis Dermatology	<b>Street Address, City, State</b>  600 Heritage Dr Ste 101, Jupiter, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced recertification survey was conducted on 9/2/20 at Genesis Dermatology, a clinical laboratory in Jupiter, Florida. Genesis Dermatology is not in compliance with 42 Code of Federal Regulations (CFR), Part 493, requirements for clinical laboratories.
<b>D6128</b>	<p><b>TECHNICAL SUPERVISOR RESPONSIBILITIES</b> CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least annually after the first year, unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individual's performance must be reevaluated to include the use of the new test methodology or instrumentation.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with laboratory personnel, there was no documentation to indicate that the technical supervisor evaluated performance of one of two testing personnel. The findings included: Review of competency records on 9/2/20 revealed that there were competencies documented for the histotechnician in 2018 and 2020, but not 2019. During an interview with the practice administrator on 9/2/20 at 10:30 a.m., she confirmed they must have missed performing the competency in 2019.</p>