Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0957961	(X3) Date Survey Completed 03/23/2021
Name of Provider or Supplier Edward A Gross Md Pa D/B/A	**Street Address, City, State** 333 Arthur Godfrey Rd Ste 302, Miami Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey conducted on 03/23/2021 found that the EDWARD A GROSS MD PA D/B/A SOUTH FLORIDA SURGICAL DERMATOLOGY clinical laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on observation, record review and testing personnel A (TP A) interview, the laboratory failed to use Eosin Y Stain Solution 1 % that was not expired for Hematoxylin and Eosin stains from 6/6/2020 to 3/23/2021. Findings Include: During laboratory tour on 03/23/2021 at 9:30 am, found in a fire safety cabinet was a 1 liter bottle with approximately 400 ml of Eosin Y Stain Solution 1 %, with lot number 1815019 that expired on 06/05/2020. Patient logs review revealed that the laboratory performed Hematoxylin and Eosin stains from 6/6/2020 to 3/23/2021 on 58 patients. During an interview on 03/23/2021 at 10:00 am, the TP A confirmed the laboratory had used expired Eosin Y Stain Solution for Hematoxylin and Eosin stains during the period of reference.