

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0958608	(X3) Date Survey Completed 09/13/2018
Name of Provider or Supplier Dermatology Group Of Florida, Pa	Street Address, City, State 5415 Park Central Ct, Naples, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory personnel, the laboratory had expired reagents in use in the laboratory. Findings include: The surveyor observed at 12:00 p.m. on 09/13/18 that the one bottle of eosin and the one bottle of hematoxylin that were in the flammable cabinet had expired. Lot # 050589 or Gill 3 Hematoxylin expired 06/18 and lot # 41312 expired on 09/18. During an interview with the office manager at 12:45 p.m., she confirmed that the reagents were expired.</p>