Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	10D0963925	04/10/2025
Name of Provider or Supplier	**Street Address, City, State**	
Forefront Dermatology, Sc Dba Skin Center,	6550 N Federal Hwy Ste 320, Fort Lauderdale, FL	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at FOREFRONT DERMATOLOGY, SC DBA SKIN CENTER on April 10, 2025. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D5413	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b) (b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports. This STANDARD is not met as evidenced by: Based on lack of documentation and interview with the laboratory technician, the laboratory did not maintain records of documentation for the Room Temperature and Humidity for 3 months (October, November and December) in 2023, out of the 24 months (April 2023 to April 2025) reviewed. Findings included: 1- Review of the ROOM TEMPERATURE/HUMIDITY LOG(s) revealed there was no temperature and humidity documented for October, November, and December of 2023. 2- Review of the Laboratory Procedure Manual for Mohs DAILY QUALITY CONTROL stated step "3 TEMPERATURE: ROOM TEMPERATURE, SLIDE OVEN, EMBEDDING CENTER, AND TISSUE PROCESSOR TEMPERATURE RECORDS MUST BE CHECKED DAILY TO ENSURE WITHIN PROPER LIMITS. THESE RECORDS

	MUST BE KEPT FOR TWO YEARS." 3- Interview on 04/10/2025 at approximately 3:10 PM with the lead histology technician stated that the records were disposed of in error during a cleanup at end of 2024.