Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0966025	(X3) Date Survey Completed 05/31/2019
Name of Provider or Supplier All Womens Health Center Of Orlando Inc	**Street Address, City, State** 431 Maitland Ave, Altamonte Springs, FL	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2155	ABO GROUP AND D(RHO) TYPING CFR(s): 493.859(c) Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if-- (1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; (2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and (3) The laboratory participated in the previous two proficiency testing events. This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to participate in proficiency testing (PT) that resulted in a score of zero percent (0%) for the Immunohematology analyte D (Rho) Typing for the first event in 2019. Findings: Review of the American Association of Bioanalysts (AAB) records for "Q1 Nonchemistry 2019" showed the laboratory received a score of "0.0 NR" (Not Reported to AAB) for the Immunohematology analyte D(Rho) Typing. During an interview on 5/31/19 at 1:25 PM, the Administrator acknowledged that the laboratory failed to get the test results submitted on time.