

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0972210	(X3) Date Survey Completed 01/18/2022
Name of Provider or Supplier Rsb Dermatology Inc	Street Address, City, State 1500 E Hillsboro Blvd Ste 204, Deerfield Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was completed on 1/18/2022 at RSB Dermatology. The laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following Conditions were not Met : 5200- General Laboratory Systems
D5200	<p>GENERAL LABORATORY SYSTEMS CFR(s): 493.1230</p> <p>Each laboratory that performs nonwaived testing must meet the applicable general laboratory systems requirements in 493.1231 through 493.1236, unless HHS approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing. The laboratory must monitor and evaluate the overall quality of the general laboratory systems and correct identified problems specified in 493.1239 for each specialty and subspecialty of testing performed.</p> <p>This CONDITION is not met as evidenced by: Based on record review and interview, the laboratory failed to perform peer-review twice a year for KOH and Scabies in 2020 and 2021.(See 5217)</p>
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to perform peer-review twice a year for KOH and Scabies in 2020 and 2021. Findings included: This a repeated deficiency from 1/28/2020. Review of American Association of Bioanalyst</p>

(AAB) Proficiency Testing Service revealed there was no records of peer-view performed twice a year for KOH in 2020. Review of Peer-Review records for Scabies revealed there was no documentation of peer-review performed for Scabies twice a year in 2020 and 2021. During an interview on 1/18/2022 at 2:34 PM, the office consultant confirmed peer-review had not been performed twice annually in 2020 for KOH and Scabies in 2020 and 2021.