

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0972210	(X3) Date Survey Completed 03/14/2024
Name of Provider or Supplier Rsb Dermatology Inc	Street Address, City, State 1500 E Hillsboro Blvd Ste 204, Deerfield Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on March 14, 2024. RSB Dermatology Inc clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5415	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (1) Identity and when significant, titer, strength or concentration. (2) Storage requirements. (3) Preparation and expiration dates. (4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: Based on observation, review of the procedure manual and quality control documents, and interview, the laboratory failed to document the making of the 70 %, 90 %, and 95 % alcohol solutions used in the processing and staining for biopsy and Mohs surgery tissue samples from 01/19/2022 to 03/14/2024. Findings: Examination of the Miles Tissue Tek VIP on 03/14/2024 at 9:40 PM, showed the processor used 70 %, and 90 % Alcohol. Examination of the flammable cabinet 03/14/2024 at 9:45 AM revealed it only contained bottles of 100 % Alcohol. Review of the procedure manual noted the biopsy stain line used 70 %, 90 %, and 95 % alcohol solutions. Review of the procedure manual noted the Mohs stain line used 95 % alcohol solution. Review of the laboratory's quality control documents revealed the laboratory did not have documentation of the preparation of the 70 %, 90 %, and 95 % alcohol solutions On 03 /14/2024 at 11:37 AM, Testing Personnel A stated they did not document the making of the 70 %, 90 %, and 95 % alcohol solutions.</p>