

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0974551	<b>(X3) Date Survey Completed</b>  01/24/2023
<b>Name of Provider or Supplier</b>  Advanced Dermatology And Cosmetic Surgery	<b>Street Address, City, State</b>  2229 N Commerce Pkwy Ste 210, Weston, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An intial survey conducted on 1/24/2023 found the ADVANCED DERMATOLOGY AND COSMETIC SURGERY clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories.
<b>D5413</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review and interview, the laboratory failed to monitor and document the laboratory temperature and humidity during at least the past two years.</p>
<b>D5433</b>	<p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(b)(1)</p> <p>For equipment, instruments, or test systems developed in-house, commercially available and modified by the laboratory, or maintenance and function check protocols are not provided by the manufacturer, the laboratory must establish a maintenance protocol that ensures equipment, instrument, and test system performance that is necessary for accurate and reliable test results and test result reporting. The laboratory must perform and document the maintenance activities specified in paragraph (b)(1)(i) of this section.</p>

This STANDARD is not met as evidenced by:

Based on observation, record review and interview, the laboratory failed to have annual maintenance performed on the microscope, or documentation showing that annual maintenance was performed during the year 2021.