

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0974551	<b>(X3) Date Survey Completed</b>  08/23/2024
<b>Name of Provider or Supplier</b>  Advanced Dermatology And Cosmetic Surgery	<b>Street Address, City, State</b>  2229 N Commerce Pkwy Ste 210, Weston, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A recertification survey conducted on 08/23/2024 found the ADVANCED DERMATOLOGY AND COSMETIC SURGERY clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories.
<b>D5413</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory recorded a Cryostat temperature outside of the acceptable range in one out of two testing days during August 2023 and tested patients. Findings included: 1- Review of "CRYOSTAT TEMPERATURE LOG" form, revealed that stated: "Action should be taken and documented if temperature is outside required range. Normal Ranges are -18C to -27 C." 2- Review of "CRYOSTAT TEMPERATURE LOG" for 19 consecutive months from February of 2023 until August of 2024 revealed that on the Month of August 2023, there was two testing dates: August 15th and August 29. The temperature recorded for August 29th was "-10C", which was outside of the acceptable range. That day the laboratory tested 14 patients. No corrective action found for that temperature record outside of the acceptable range. During an interview on 08/23 /2024 at 12:20 PM with office manager, she confirmed that the Cryostat Temperature recorded for that day was outside of the acceptable range and no corrective action was documented</p>