

|  |  |   |
|--|--|---|
| <b>Statement of Deficiencies</b>   | <b>(X1) Provider/Supplier/CLIA Identification Number</b><br><br>10D0978533 | <b>(X3) Date Survey Completed</b><br><br>08/22/2018 |
| <b>Name of Provider or Supplier</b><br><br>Skin And Cancer Associates Llp  | <b>Street Address, City, State</b><br><br>3275 N Sr 7, Margate, FL         |   |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. |  |   |

| <b>(X4) ID Prefix Tag</b> | <b>Summary Statement of Deficiencies</b>   |
|---------------------------|--|
| <b>D5609</b>              | <p>HISTOPATHOLOGY<br/>CFR(s): 493.1273(e)(f)</p> <p>(e) The laboratory must use acceptable terminology of a recognized system of disease nomenclature in reporting results. (f) The laboratory must document all control procedures performed, as specified in this section.</p> <p>This STANDARD is not met as evidenced by:<br/>Based on record review and interview, the laboratory failed to document the open dates for reagents used in their Hematoxylin &amp; Eosin (H&amp;E) stains from 8/22/16 to 8/22/18, and failed to retain the reagent logs for 2017. Findings: Record review of the laboratory's "Chemical Substance Log" showed that the laboratory failed to record when the reagents for their H&amp;E stains were opened, and that they did not have the reagent logs for 2017. The laboratory maintains separate reagent logs for Alcohol 100%, Hematoxylin III Gills, Eosin Y Stain, Histo Clear, and Alcohol/Histo clear. During an interview on 8/22/18 at 10 AM, Testing Personnel C acknowledged that the laboratory did not record open dates for reagents used in their H&amp;E stains, and that the reagent logs 2017 were lost.</p> |