

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0978533	(X3) Date Survey Completed 09/30/2020
Name of Provider or Supplier Skin And Cancer Associates Llp	Street Address, City, State 3275 N Sr 7, Margate, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey conducted on 9-30-2020, found that Skin and Cancer Associates LLP clinical laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D5415	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (1) Identity and when significant, titer, strength or concentration. (2) Storage requirements. (3) Preparation and expiration dates. (4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: Based on observations and interviews, the laboratory was found to have expired Eosin 1% Alcoholic reagent in the biosafety cabinet. Findings Included: Observation of Histopathology Room revealed a biosafety cabinet that had Eosin 1% Alcoholic reagent that expired on 7-17-2020. An interview on 9-30-2020 at 10:15am, with the office manager confirmed that Eosin 1% Alcoholic reagent expired on 7-17 -2020.</p>