

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D0990498	(X3) Date Survey Completed 04/07/2023
Name of Provider or Supplier Boca Raton Dermatology Pa	Street Address, City, State 5258 Linton Blvd Ste 306, Delray Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Boca Raton Dermatology PA on 04/07/2023. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review, and interview with the Laboratory Director, the laboratory failed to follow manufacturer's instructions to store flammable Reserve Eosin Multichrome in a flammable area for two out of two years (2021 - 2023). Findings Included: A tour of the laboratory on 04/07/2023 at 9:10 AM revealed two containers of Reserve Eosin Multichrome (Lot number 144955) on a cart. Review of the manufacturer's label on the Reserve Eosin Multichrome revealed this was a flammable liquid and vapor and required storage in a fireproof area. Review of the laboratory's procedure "Chemical Storage" revealed "Flammable liquids are stored in flammable storage cabinets with self-closing doors and proper ventilation..." Review of the Material Safety Data Sheet for Eosin Multichrome Reserve revealed "Store in flammable liquid safety cabinet when possible." On 04/07/23 at 11:50 AM, the Laboratory Director stated he did not know the Reserve Eosin Multichrome needed to be stored in a flammable cabinet.</p>