

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D0998636	<b>(X3) Date Survey Completed</b>  01/27/2020
<b>Name of Provider or Supplier</b>  University Of Miami Hospital And Clinics -	<b>Street Address, City, State</b>  8170 Royal Palm Blvd, Coral Springs, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A recertification survey conducted on 01/27/2020 found that University of Miami Hospital and Clinics Coral Springs Clinical Laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories.
<b>D2121</b>	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of College of American Pathologists (CAP) proficiency testing records and interview with Technical Supervisor (TS), the laboratory failed to score at least 80% on White Blood Cell Identification (WBCI) for 1 (3rd event of 2018) out of 6 events for Hematology specialty reviewed during 2018 and 2019. Findings include: Review of CAP proficiency testing records revealed a score of 0% for WBCI in the 3rd event of 2018. During an interview on 01/27/2020 at 11:30 AM, the TS confirmed the failed proficiency testing score.</p>