

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1001532	(X3) Date Survey Completed 06/15/2026
Name of Provider or Supplier University Park Dermatology	Street Address, City, State 8451 Shade Ave Ste 205, Sarasota, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at University Park Dermatology on 06/15/2026. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiencies are cited as follows:
D3011	FACILITIES CFR(s): 493.1101(d) Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials. This STANDARD is not met as evidenced by: Based on observation, interview, and review safety data sheets (SDS), the laboratory failed to ensure protection from chemical hazards used for Histopathology testing of Hematoxylin and Eosin (H&E) stain from 04/04/2024 to 06/15/2026. Findings: 1. During a tour of the laboratory on 06/15/2026 at 10:20 AM, signage for Bluing Reagent, Eosin Y, Alcohol, Reagent 100%, and Safe-Clear all included directions to avoid breathing fumes. These reagents are used for the staining of H & E slides. No fume hood was observed to provide protection from respiratory chemical hazards. 2. The SDS sheets for Bluing Reagent, Eosin Y, Alcohol, Reagent 100%, and Safe-Clear all included hazard statements regarding inhalation precautions and to not breathe "dust/ fume/ gas/ mist/vapors/spray". 3. The Histology Tech and Practice Administrator both confirmed on 06/15/2026 at 10:25 AM there was no hood or any other method to ensure respiratory protection from chemical hazards used for H & E staining.