

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1002529	(X3) Date Survey Completed 04/08/2026
Name of Provider or Supplier Aqua Dermatology Of Florida Pa	Street Address, City, State 21550 Angela Lane, Venice, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Aqua Dermatology of Florida dba Coastal Dermatology on 4/8/26. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiencies cited are as follows:
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review, and interview, the laboratory failed to monitor and document two of two cryostat temperatures for Histopathology testing. Finding included: 1. On 4/8/26 at 11:05 a.m. two cryostat (serial numbers 02.1999 and 04.2006) used for Histopathology testing were observed powered up in the laboratory with signage to not turn off the instruments. 2. The MOHS Manual was last reviewed by the Laboratory Director on 1/5/26. The MOHS Manual included a MOHS Cryotomy Procedure which directed under Quality Control #2-cryostat temperature to be between -20 to -30 degrees Celsius to be checked and logged every day of testing. 3. Three of three months (3/2026, 10/2025, and 7/2024) of Daily QC (Quality Control) worksheets reviewed included space to document temperature for one cryostat, there was no indication of which of two cryostats temperature was</p>

documented daily. 4. The Clinical Supervisor on 4/8/2026 at 12:05 p.m. verified the laboratory used both cryostats but did not have documentation of monitoring temperatures for both each day used for testing.