Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 10D1010469	**(X3) Date Survey Completed** 06/28/2024
Name of Provider or Supplier Monica K Bedi Md Pa	**Street Address, City, State** 3830 Bee Ridge Rd Ste 200, Sarasota, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	An announced CLIA recertification survey was conducted at Monica K Bedi MD PA dba Dermatology Associates on 06/28/2024. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiency:
D5791	ANALYTIC SYSTEMS QUALITY ASSESSMENT CFR(s): 493.1289(a)(c) (a) The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and when indicated, correct problems identified in the analytic systems specified in 493.1251 through 493.1283. (c) The laboratory must document all analytic systems assessment activities. This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to follow written policies and procedures for an ongoing mechanism to monitor, assess, and correct problems in the analytic systems for 11 of 12 months reviewed for 2023. Findings included: 1. Review of the Quality Assurance (QA) program approved by the Laboratory Director on 10/24/2023, showed Monthly QC [Quality Control] - This form should be completed at the end of each month and signed by the lab director..." 2. Review of the Monthly QC Worksheets for 2023 revealed 11 (January, February, March, April, May, June, July, September, October, November, and December) of the 12 months were signed and dated by the Laboratory Director prior to the worksheets being filled out. There was no documentation of QC on the worksheets. 3. The Chief Administrative Office confirmed on 6/28/2024 at 10:50 AM the laboratory had not followed the written QA policy for 11 of 12 months in 2023.