

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1017753	(X3) Date Survey Completed 03/18/2022
Name of Provider or Supplier Advanced Dermatology And Cosmetic Surgery	Street Address, City, State 1617 Tamiami Trl, Port Charlotte, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced recertification survey was conducted 3/11/22 through 3/18/22 at Advanced Dermatology and Cosmetic Surgery, a clinical laboratory in Port Charlotte, Florida. Advanced Dermatology and Cosmetic Surgery is not in compliance with Code of Federal Regulations (CFR) 42, Part 493, Laboratory Requirements. The following is a description of the noncompliance.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview the Laboratory failed to verify the accuracy of the Histopathology MOHS testing at least twice a year for 1 (2020) out of 2 years (2020-2021) reviewed. Findings Included: Review of the policy and procedures (last revision signed by the Laboratory Director on 02/02/2021) revealed that under "Quality Assurance Manual" it stated that the Laboratory Director's duty was to "Ensure that the laboratory is enrolled in a HCFA-approved proficiency-testing program." On 03/11/2022 at 2:00 p.m., the Office Manager confirmed that the laboratory was using peer review to verify the accuracy of Histopathology MOHS testing. Review of Peer Review records for 2020 revealed that testing was only performed on 08/10/2020. On 03/18/2022 at 1:36 p.m., the Office Manager confirmed via email that the peer review for Histopathology MOHS testing only occurred once in 2020.</p>