

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1018194	(X3) Date Survey Completed 11/26/2019
Name of Provider or Supplier Ecc Medical Group	Street Address, City, State 8335 Nw 12th St, Miami, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA validation survey conducted at FM Antonietta Schettino MD PA /dba/ ECC Pediatrics on 11/26/2019, found the facility was not in compliance with 42 CFR Part 493 CLIA requirements.
D2121	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency testing (PT) results and interview with laboratory director (LD), the laboratory failed to score at least 80 % on Platelets analyte for 1 (3rd event) out of 5 events reviewed in 2018 and 2019. Findings include: Review of American Proficiency Institute PT records for 2018 and 2019, revealed a score of 40 % for Platelets in the 3rd event of 2018. During an interview on 11/26/2019 at 10:30 am, the LD confirmed the failed PT score for the analyte of reference.</p>