Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1033336	(X3) Date Survey Completed 06/06/2018
Name of Provider or Supplier Uf Health Oncology-Baymeadows	**Street Address, City, State** 7751 Bay Meadows Rd Ste 205, Jacksonville, FL	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2010	TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(2) The laboratory must test samples the same number of times that it routinely tests patient samples. This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to test hematology proficiency testing samples the same number of times that it tests patient samples for one proficiency testing event in 2017. The findings include: A record review of 2018 American Proficiency Institute (API) proficiency testing was done on June 6, 2018. The documentation showed testing of the five proficiency samples was performed twice on March 27th, 2017. The tests performed were: White Cell Count, Red Cell Count, Hemoglobin, Hematocrit, Platelet Count, Mean corpuscular volume (MCV), Red cell distribution width (RDW), Neutrophils, Lymphocytes, Monocytes, Eosinophils, and Basophils. CTR-01: 3/27/17 at 10:19 and 10:20 CTR-02: 3/27/17 at 10:22 and 10:23 CTR-03: 3/27/17 at 10:24 and 10:26 CTR-04: 3/27/17 at 10:28 and 10:29 CTR-05: 3/27/17 at 10:30 and 10:33 Interview with the technical consultant on 6 /6/18 at 9:27am confirmed hematology proficiency testing had been performed in duplicate against laboratory policy.