

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D1033336	<b>(X3) Date Survey Completed</b>  08/30/2019
<b>Name of Provider or Supplier</b>  Uf Health Oncology-Baymeadows	<b>Street Address, City, State</b>  7751 Bay Meadows Rd Ste 205, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	At the time of the announced, onsite recertification survey, UF Health Oncology Baymeadows was found to not be in compliance with the CLIA laboratory requirements of 42 CFR 493.
<b>D2121</b>	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of American Proficiency Institute (API) proficiency testing results and staff interview, the laboratory had 1 out of 3 testing events in 2018 that had an analyte that received less than 80% in the specialty of Hematology. The findings include: Review of API proficiency testing results found the 2nd testing event in 2018 had a 60% for White Blood Cell Differential. During an interview on 8/30/19 at 10:20am, the Technical Consultant confirmed the proficiency testing failure stating it was caused by the instrument being in the wrong mode during testing.</p>