

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D1042584	<b>(X3) Date Survey Completed</b>  06/11/2025
<b>Name of Provider or Supplier</b>  Bradley P Grant Md Pa	<b>Street Address, City, State</b>  1480 S Wickham Rd, West Melbourne, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at Bradley P Grant on June 11, 2025. The laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on interview and review of chemical and surgical case logs, the laboratory failed to use non-expired reagents in their Hematoxylin and Eosin stain for staining patient's histopathology slides from 04/04/2023 to 04/11/2023 and 06/04/2024 to 06/18/2024. Findings: 1. Review of the Chemical Expiration Dates log for April 2023 listed the Clarifier lot #104693 expired on 3/2023, and Clarifier lot #125131 expired on 11/2024 and was opened on 06/19/2024. 2. Review of the Mohs Surgical Case Log revealed there were seven surgical procedures on 04/04/2023, and seven surgical procedures on 04/11/2023. 3. Review of the Chemical Expiration Dates log for June 2024 listed the Clear-Rite lot #2015108 expired on 06/02/2024, and Clear-Rite Lot #2412323 expiring on 05/06/2028 was opened on 06/19/2024. 4. Review of the Mohs Surgical Case Log revealed there were seven surgical procedures on 06/04/2024, seven surgical procedures on 06/11/2024, and seven surgical procedures on 06/18/2024. 5. During an interview on 06/11/2025 at 11:45 AM, the Mohs Technician confirmed the Clear-Rite and Clarify were on back order and expired reagents were used.</p>