

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D1046775	<b>(X3) Date Survey Completed</b>  08/13/2025
<b>Name of Provider or Supplier</b>  Dermatology Associates Of Palm Beaches Pllc	<b>Street Address, City, State</b>  2605 W Atlantic Ave Ste D-204, Delray Beach, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at Dermatology Associates of Palm Beaches PLLC on 8/13/2025. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
<b>D3011</b>	<p><b>FACILITIES</b> CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on observation, record review, and interview it was determined the laboratory failed to ensure protection from chemical hazards. Findings include: 1. On 8/13/25 at 9:20 AM, Plantinumline 95% Reagent Alcohol lot 2219211 expiration date 1/11/2025 was observed stored in the cabinet under the sink instead of in the flammable safety cabinet. There was signage on the wall for all reagents to be stored in the flammable cabinet. 2. The laboratory failed to have a fume hood or air quality monitoring device to ensure protection from chemical hazardous fumes from the reagents (95% alcohol, Pro-Par Clearant, and Hematoxylin) used by the laboratory which had hazard labels for respiratory health hazard and directed to not breathe mist/vapors/spray. 3. The Risk Manager at 9:20 AM, stated the 95% Alcohol was to be stored in the flammable safety cabinet and that the laboratory was not aware protection from chemical hazardous fumes was required.</p>