

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1048255	(X3) Date Survey Completed 05/09/2023
Name of Provider or Supplier Advanced Gastroenterology And Surgery Associates	Street Address, City, State 26540 Ace Ave, Ste A, Leesburg, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on May 9, 2023. Advanced Gastroenterology and Surgery Associates clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5609	<p>HISTOPATHOLOGY CFR(s): 493.1273(e)(f)</p> <p>(e) The laboratory must use acceptable terminology of a recognized system of disease nomenclature in reporting results. (f) The laboratory must document all control procedures performed, as specified in this section.</p> <p>This STANDARD is not met as evidenced by: Based on review of the reagent log and interview, the laboratory failed to record the open dates for all reagents used in their Hematoxylin & Eosin (H & E), Alcain Blue /Periodic Acid-Schiff (AB/PAS), Diff Quick, and Trichrome stains from 08/02/2021 to 05/09/2023. Findings: The laboratory used the following reagents: Acetic Alcohol, Alcain Blue, Bluing Reagent, Diff Quik, Eosin, Hematoxylin, Periodic Acid, Schiffs Reagent, and Xylene. Review of the laboratory's "Reagent Lot Number and Expiration" log showed there were no open dates listed. On 05/09/2023 at 10:35 AM, the Testing Personnel acknowledged the open dates were not recorded.</p>