Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1066947	(X3) Date Survey Completed 08/29/2018
Name of Provider or Supplier Center Of Orlando For Women Llc	**Street Address, City, State** 1103 Lucerne Terrace, Orlando, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2153	ABO GROUP AND D(RHO) TYPING CFR(s): 493.859(a) Failure to attain a score of at least 100 percent of acceptable responses for each analyte or test in each testing event is unsatisfactory analyte performance for the testing event. This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to get a 100% on proficiency testing for the 3rd testing event in 2017. Findings: Review of American Association of Bioanalysts (AAB) proficiency testing records showed that the laboratory received a score of 80% for D (Rho) Typing for the 3rd testing event in 2017. During an interview on 8/29/18 at 11:00 AM, the Administrator acknowledged the proficiency testing failure.