

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D1079582	<b>(X3) Date Survey Completed</b>  03/25/2024
<b>Name of Provider or Supplier</b>  Academic Alliance In Dermatology Inc	<b>Street Address, City, State</b>  2044 Trinity Oaks Blvd Ste 2130, Trinity, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An on - site announced CLIA recertification survey was conducted at Academic Alliance in Dermatology Inc. on 03/20/2024 - 03/25/2024. The laboratory is not in compliance with Code of Federal Regulations (CFR) 42, Part 493, Laboratory Requirements. The following is a description of the standard level deficiencies:
<b>D5217</b>	<p><b>EVALUATION OF PROFICIENCY TESTING PERFORMANCE</b> CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Laboratory Manager, the laboratory failed to ensure the accuracy for twice a year testing for parasitology (scabies) testing for Testing Personnel #A for two out of two years reviewed (2022-2023). The findings included: A record review of the laboratory's policy titled "Potassium Hydroxide (KOH) Examination of Skin Hair, or Nails" revealed, "Quality control analysis must be performed every day that patient samples are tested." A record review of the "Proficiency Testing Log Sheet KOH &amp; Scabies Log" revealed that Testing Personnel #A had only verified accuracy for scabies on 7/24/23 for two out of two years reviewed (2022 - 2023). On 03/27/24 at 10:15 AM, the Laboratory Manager confirmed that there was missing semi-annual verification of scabies testing.</p>