

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1081549	(X3) Date Survey Completed 02/09/2024
Name of Provider or Supplier Sunset Pediatrics Llc	Street Address, City, State 7300 Sw 62nd Place Ph-West, South Miami, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An on - site announced CLIA recertification survey was conducted at Sunset Pediatrics. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D5313	<p>SPECIMEN SUBMISSION, HANDLING, AND REFERRAL CFR(s): 493.1242(b)</p> <p>The laboratory must document the date and time it receives a specimen.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Lab Director, the laboratory failed to document the storage temperature for Boule Con-Diff Tri-Level Hematology Control material for 2 of 2 years (2022-2023) reviewed. Findings include: 1. Review of laboratory temperature logs on 2/9/2024 at 11:00 a.m., revealed there were no temperatures documented for the refrigerator where the Boule Con-Diff Tri-Level Hematology Control material was stored. The Manufacturer listed the acceptable storage temperature range as 2-10 degrees Celsius. 2. The Lab Director confirmed on 2 /9/2024 at 11:20 a.m. the laboratory had not documented the temperature for the refrigerator where the Boule Con-Diff Tri-Level Hematology Control material was stored for 2 of 2 years (2022-2023) reviewed.</p>