

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1082767	(X3) Date Survey Completed 08/26/2022
Name of Provider or Supplier Ameripath Florida Llc	Street Address, City, State 6806 N Sr 7, Coconut Creek, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey conducted on 08/26/2022 found the AMERIPATH FLORIDA LLC clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D3031	<p>RETENTION REQUIREMENTS CFR(s): 493.1105(a)(3)</p> <p>Analytic systems records. Retain quality control and patient test records (including instrument printouts, if applicable) and records documenting all analytic systems activities specified in 493.1252 through 493.1289 for at least 2 years.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to retain stain process records for year 2021. Findings include: Review of laboratory's quality control documents revealed the laboratory failed to have the following records for year 2021: - Min/Max Humidity Monitoring Log. -Room temperature log. -Cryostat temperature. - Patient log. -Reagent Acceptability Worksheet. -Stain/Technical Quality. -Equipment maintenance. During an interview on 08/26/2022 at 10:30 AM, the General Supervisor confirmed that the laboratory failed to keep the documents of reference for year 2021.</p>