

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1083698	(X3) Date Survey Completed 01/23/2020
Name of Provider or Supplier Neuropain Center,The	Street Address, City, State 1931 Martin Luther King Blvd, Panama City, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at The Neuropain Center on January 23, 2020. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the supervisor, the laboratory failed to perform competency evaluations on 2 out of 3 clinical consultants for 2 of 2 years (2018-2019) reviewed. Findings Included: Review of the CMS 209 revealed 3 clinical consultants. Review of the documents revealed no competency evaluations for 2 of 3 clinical consultants from January 2018 through December 2019. During the interview on 1/23/20 at 10:45 AM , the laboratory supervisor confirmed that there was no documented competency evaluations for clinical consultants # 4 and # 5.</p>