

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1087175	(X3) Date Survey Completed 07/03/2018
Name of Provider or Supplier Naples Center For Dermatology & Cosmetic Surgery	Street Address, City, State 9980 Central Park Blvd N #124, Boca Raton, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5429	<p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(1)</p> <p>For unmodified manufacturer's equipment, instruments, or test systems, the laboratory must perform and document maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with testing person, the laboratory failed to perform and record microscope maintenance for the year 2017 from two-year review period (2016 - 2018) for the subspecialty of histopathology. The findings include: On 7/3/18 at 11:30AM, instrument maintenance records from January 1, 2016 to July 3, 2018 showed documentation of microscope maintenance for dates 2/25/2016 and 6 /2018, no records for due date of 2/2017 for the year 2017. During an interview on 7/3 /18, at 12:30 PM, the testing person confirmed that the laboratory did not perform yearly microscope maintenance for the due date of 2/2017 for the year 2017.</p>