

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D1087510	<b>(X3) Date Survey Completed</b>  03/07/2018
<b>Name of Provider or Supplier</b>  Suncoast Skin Solutions Inc	<b>Street Address, City, State</b>  11200 Seminole Blvd Ste 205, Largo, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5209</b>	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based upon record review and interview with the Laboratory Manager, the laboratory failed to perform a competency evaluation for 1 out of 1 (#A) testing personnel for 2 out of 2 years (2016-2018). Findings included: At the time of survey, competency records for Testing Personnel #A were requested but the laboratory was unable to provide the competency records for 2 out of 2 years (2016-2018). The procedure manual for the laboratory did not contain a procedure for evaluating competency of testing personnel. During the interview on 03/07/18 at 12:30 p.m., the Laboratory Manager confirmed the laboratory did not have a competency procedure for the survey time period ( 2016-2018) and did not perform the competency evaluations for Testing Personnel #A.</p>