

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1094272	(X3) Date Survey Completed 07/15/2020
Name of Provider or Supplier Aqua Dermatology Of Florida Pa	Street Address, City, State 4180 Us 27 South, Sebring, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Waters Edge Dermatology LLC on 07/15/20. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Lab Coordinator, the laboratory failed to evaluate the accuracy of the subspecialty parasitology Scabies testing at least twice a year for 2 out of 2 years (2018-2020) reviewed. Findings Included: Record review of the laboratory's policy and procedure revised 1/15/2014 for "Verification Testing KOH Prep" revealed the laboratory "will perform a 6 month verification of KOH prep reading at 6 month intervals for every provider for all sub-specialties...If there is no sub-specialties testing to be performed during this period for specific locations then testing materials will be brought in from another location, so that verification testing for each provider can be performed". Review of records for Scabies testing revealed no documentation for the verification of accuracy was available for 2018-2020. Interview on 07/15/20 at 11:20 a.m. with the Lab Coordinator, confirmed that verification of accuracy records for Scabies testing were not available for review. She also stated she did not know that the 6 month verification of accuracy had to be performed even if patients were not tested.</p>