

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  10D1094392	<b>(X3) Date Survey Completed</b>  03/13/2019
<b>Name of Provider or Supplier</b>  Skin And Cancer Associates Llp	<b>Street Address, City, State</b>  17170 Royal Palm Blvd, Weston, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5609</b>	<p><b>HISTOPATHOLOGY</b> CFR(s): 493.1273(e)(f)</p> <p>(e) The laboratory must use acceptable terminology of a recognized system of disease nomenclature in reporting results. (f) The laboratory must document all control procedures performed, as specified in this section.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to provide complete histopathology quality control documentation from 3/13/17 to 3/13/19. Findings: Record review of quality control documentation showed there was no records of a reagent log with the lot numbers, expiration dates and open dates for the reagents used in their Hematoxylin &amp; Eosin (H &amp; E) stains. During an interview on 3/13/19 at 10:15 AM, the Risk Manager stated that he was unable to find a reagent log.</p>