

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D1096369	(X3) Date Survey Completed 12/24/2025
Name of Provider or Supplier George S Sidhom Md Pa D/B/A Brandon Pain Clinic	Street Address, City, State 722 Bowing Oak Dr, Brandon, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at George S Sidhom MD PA dba Brandon Pain Clinic on 12/23/25 through 12/24/25. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Standard deficiencies cited are as follows:
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>(b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3)-- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to perform and document calibration verification at least once every 6 months for the speciality of Chemistry and Toxicology performed from 2/13/24 to 12/23/2025. Findings included: 1. Calibration Verification reports for speciality of Chemistry and Toxicology which</p>

used two point calibration performed from 2/13/24 to 12/23/2025 was requested. No documentation was presented for review. 2. The Technical Supervisor on 12/23/2025 at 11:11 a.m. confirmed the laboratory had not performed and documented calibration verification from 2/13/24 to 12/23/2025 every 6 months for the speciality of Chemistry and Toxicology.

D6080

LABORATORY DIRECTOR RESPONSIBILITIES

CFR(s): 493.1445(c)

(c) The laboratory director must: (c)(1) Be onsite at least once every 6 months, with at least 4 months between the minimum two on-site visits. Laboratory directors may elect to be on-site more frequently and must continue to be accessible to the laboratory to provide telephone or electronic consultation as needed; and (c)(2) Provide documentation of these visits, including evidence of performing activities that are part of the laboratory director responsibilities.

This STANDARD is not met as evidenced by:

Based on record review and interview, the Laboratory Director failed to establish a policy to be onsite once every six months and failed to document any onsite visits from 1/1/2025 to 12/23/2025. Findings included: 1. The Laboratory Policy and Procedure signed by the Laboratory Director on 7/07/2020, failed to contain a policy to be onsite once every six months and to document any onsite visits. 2. On 12/23/2025 at 11:31 a.m., the Technical Supervisor confirmed the laboratory failed to have an established policy for the Laboratory Director to be onsite once every six months and failed to have documentation of any onsite visits from 1/1/2025 to 12/04/2025.