Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 10D1100590	**(X3) Date Survey Completed** 07/05/2023
Name of Provider or Supplier Dermatology Consultants Pa	**Street Address, City, State** 12600 Pembroke Rd Ste 310, Miramar, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	A recertification survey conducted on 07/05/2023 found the DERMATOLOGY CONSULTANTS PA clinical laboratory not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D5413	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports. This STANDARD is not met as evidenced by: Based on record review and staff interview, the laboratory failed to document the cryostat temperature in one testing day and recorded a temperature outside of the acceptable range in one testing day for the period reviewed of July 2021 to June 2023. Findings include: Review of the "Daily QC Worksheet" form revealed that the "acceptable cryostat temperature range: -14C to -30C". Review of Daily QC records from 07/01/2021 to 06/30/2023 revealed the following for the cryostat 1510S working temperatures: -No record of the cryostat temperature on 09/26/2022, 10 patients were tested on that day. -Recorded a temperature of 73C which is outside of the acceptable range on 02/06/2023, 13 patients were tested on that day. During an interview on 07 /05/2023 at 10:45 AM, the laboratory consultant confirmed that the laboratory had the incidences with the cryostat temperature records on the days listed above.