

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2012469	(X3) Date Survey Completed 04/02/2026
Name of Provider or Supplier Jacksonville Beach Pediatric Care Center	Street Address, City, State 8990 Rg Skinner Pkwy, Jacksonville, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Privia Medical Group Florida Llc on 4/3/2026. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D2010	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(2)</p> <p>(b)(2) The laboratory must test samples the same number of times that it routinely tests patient samples.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to ensure that proficiency testing (PT) samples were tested in the same manner as patient specimens for 5 of 5 samples reviewed (#6 through #10) during the 2025 second hematology testing event. Findings include: 1. A review of the American Proficiency Institute (API) hematology proficiency testing records on 4/2/2026 revealed that during the second testing event of 2025, PT samples were run in duplicate. The instrument printouts dated 7/29/2025 showed the following duplicate run times for the hematology challenges: Sample #6: Run at 15:27 and 15:28 Sample #7: Run at 15:30 and 15:32 Sample #8: Run at 15:33 and 15:35 Sample #9: Run at 15:59 and 16:01 Sample #10: Run at 16:02 and 16:03 2. The laboratory 's PT attestation statement for this event was signed by Testing Person #5 and Testing Person #1 on 7/29/2025, and by the Laboratory Director on 8/26/2025. The statement signed by all three individuals certified that "as closely as possible, these proficiency testing samples were tested in the same manner as patient specimens." 3. In an interview on 4/2/2026 at 11:00 AM, Testing Person #4 confirmed that the laboratory does not routinely run patient specimens in duplicate and agreed that the 2025 second event hematology PT samples should not have been run in duplicate.</p>