

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2021644	(X3) Date Survey Completed 06/26/2018
Name of Provider or Supplier Ciocca Dermatology Pa	Street Address, City, State 7001 Sw 97th Ave Ste 101, Miami, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review of Wisconsin State Laboratory of Hygiene Proficiency Testing (WSLH PT) and interview with laboratory personnel, the laboratory did not verify the accuracy of fungal cultures twice annually. Findings include: Review of 2 years of WSLH PT revealed that the laboratory failed to submit results for the first event of 2018. During an interview on 6/25/2018 at 10:30 AM, the Office Manager confirmed that the laboratory failed to submit the results for the event of reference.</p>