

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2029711	(X3) Date Survey Completed 05/15/2019
Name of Provider or Supplier Vipul Joshi Md Pa D/B/A	Street Address, City, State 1355 Providence Rd, Brandon, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Vipul Joshi MD PA DBA Bay Area Arthritis & Osteoporosis on 05/15/19. The laboratory is not in compliance with 42 CFR Part 493, Requirements for Laboratories. The following is a description of the standard level deficiencies:
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Office Manager the laboratory failed to perform competency assessments on the Technical Supervisor and Testing Person for 1 out of 1 (2018-2019) year. Findings Included: Review of the CMS 209, Laboratory Personnel Report, dated and signed by the Laboratory Director on 05/08 /19 revealed the Technical Supervisor was also the Testing Person. A review of the Technical Supervisor's personnel file revealed a 07/10/18 date of hire. No competencies were completed on his ability as the Technical Supervisor or competency on him being the Testing Person since his hire date. During an interview on 05/15/19 at 2:30 PM the Office Manager confirmed that there were no competency evaluations performed on the Technical Supervisor/Testing Person.</p>