

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2036991	(X3) Date Survey Completed 01/15/2020
Name of Provider or Supplier University Of Miami Hospital And Clinics -	Street Address, City, State 8100 Sw 10th Street Crossroad Business Park Big 3, Plantation, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey conducted on 01/15/2020 found that the University of Miami Hospital and Clinics Plantation clinical laboratory was not in compliance with 42 CFR Part 493, Requirements for Laboratories.
D2087	<p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review of College of American Pathologists (CAP) proficiency testing records and interview with General Supervisor (GS), the laboratory failed to score at least 80 % on Total Protein for 1 (3rd event of 2018) out of 6 events for Routine Chemistry reviewed. Findings include: Review of CAP proficiency records revealed a score of 20 % for Total Protein in the 3rd event of 2018. During an interview on 01/15/2020 at 11:30 AM, the GS confirmed the failed proficiency testing score.</p>