

| | | |
|--|---|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 10D2037541 | (X3) Date Survey Completed 11/27/2018 |
| Name of Provider or Supplier Bay Dermatology & Cosmetic Surgery Pa | Street Address, City, State 7739 S Suncoast Blvd, Homosassa, FL | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D6102 | <p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(12)</p> <p>The laboratory director must ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the Assistant Office Manager, the laboratory director did not ensure that there was documentation to indicate that the two out of four testing personnel (#C and #D) had laboratory training to perform the mycology, parasitology, and histology testing. Findings included: Review of testing personnel records for Testing Personnel #C and #D revealed that the laboratory did not have training records for #C and #D for mycology, parasitology, and histology testing. Interview at 2:30 PM on November 27, 2018 with the Assistant Office Manager confirmed there were no training records for #C and #D for mycology, parasitology, and histology.</p> |