

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 10D2037564	<b>(X3) Date Survey Completed</b> 02/27/2024
<b>Name of Provider or Supplier</b> Skin And Cancer Associates Llp	<b>Street Address, City, State</b> 1447 Medical Park Blvd Ste 107, Wellington, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A recertification survey was conducted on February 27, 2024. SKIN AND CANCER ASSOCIATES LLP clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of peer review records and interview, the laboratory failed to maintain documentation to verify accuracy of the reading and interpretation of the Hematoxylin and Eosin (H&amp;E) stain at least twice annually for 2022 and 2023. Findings: Review of peer review records revealed the peer review was done on 09/09 /2022 and 10/06/2023. On 02/27/2024 at 10:57 AM, the Laboratory Consultant stated peer review was only done once each year.</p>