

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 10D2045091	(X3) Date Survey Completed 04/06/2022
Name of Provider or Supplier Ivf Life Inc	Street Address, City, State 1912 Boothe Circle, Second Floor, Longwood, FL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A recertification survey was conducted on April 6, 2022. IVF Life Inc clinical laboratory was not in compliance with 42 CFR 493, requirements for clinical laboratories.
D5805	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to have the correct address of the laboratory for four (#2, #3, #5, #6) of six (#1, #2, #3, #4 #5, #6) patients' laboratory test reports and failed to list the normal values of the laboratory tests for four (#3,#4, #5, #6) of six (#1, #2, #3, #4 #5, #6) patients' laboratory test reports. Findings: 1. Review of the patients' laboratory test reports showed the address on the reports of patients #2, #3, #5, and #6, was the address of the laboratory's previous location. On 04/06/2022 at 2:45 PM, the Laboratory Director stated the laboratory moved to a new location in August or September and they needed to change the address on the reports to the current location. 2. Review of the patients' laboratory test reports showed the endocrinology normal values for patients #3, #4, #5, and #6, were not listed on the laboratory reports. On 04/06/2022 at 2:40 PM, the Laboratory Director said the endocrinology normal values were not on the laboratory reports.</p>